



Dave Heineman  
Governor

# State of Nebraska

DEPARTMENT OF AGRICULTURE  
GREGORY A. IBACH  
Director

Nebraska Standards Laboratory  
3721 West Cuming Street  
Lincoln, NE 68524

## CERTIFICATION FOR STANDARDS OF MEASUREMENT

Agency: Food Safety & Consumer Protection  
Area 35

Date: 11/5/2012  
Agency ID: 35  
NE Cert No: 20120130  
P.O. No:

Submitted By: Stoney Brooks

Electronic Copy Page 1

Denomination	Serial Number	Weight Number	Tolerance	Accept/Reject
50 lb	A5C-1	1	F	A
50 lb	A5C-11	1	F	A
50 lb	A5C-4	1	F	A
1000 lb	B10	1	F	A
1000 lb	B11	1	F	A
1000 lb	B12	1	F	A
1000 lb	B13	1	F	A
1000 lb	B14	1	F	A
1000 lb	B17	1	F	A
1000 lb	B18	1	F	A
1000 lb	B19	1	F	A
1000 lb	B2	1	F	A
1000 lb	B20	1	F	A
1000 lb	B21	1	F	A
1000 lb	B22	1	F	A
1000 lb	B23	1	F	A
1000 lb	B3	1	F	A
1000 lb	B4	1	F	A
1000 lb	B5	1	F	A
1000 lb	B6	1	F	A
1000 lb	B7	1	F	A
1000 lb	B8	1	F	A
1000 lb	B9	1	F	A
50 lb	B-C-1	1	F	A
50 lb	B-C-10	1	F	A

Department of Agriculture, 301 Centennial Mall South  
Web Site: [www.agr.ne.gov](http://www.agr.ne.gov)

Administration  
P.O. Box 94947  
Lincoln, NE 68509-4947  
(402) 471-2341  
FAX: (402) 471-6876

Agriculture Laboratories  
3703 South 14<sup>th</sup> Street  
Lincoln, NE 68502-5399  
(402) 471-2176  
FAX: (402) 471-0091

Bureau of Animal Industry  
P.O. Box 94987  
Lincoln, NE 68509-4787  
(402) 471-2351  
FAX: (402) 471-6893

Bureau of Dairies & Foods  
P.O. Box 95064  
Lincoln, NE 68509-5064  
(402) 471-2536  
FAX: (402) 471-2759

Bureau of Plant Industry  
P.O. Box 94956  
Lincoln, NE 68509-4756  
(402) 471-2394  
FAX: (402) 471-6892

Weights and Measures  
P.O. Box 94757  
Lincoln, NE 68509-4757  
(402) 471-4292  
FAX: (402) 471-2759

Submitted By: Stoney Brooks

Electronic Copy Page 2

<u>Denomination</u>	<u>Serial Number</u>	<u>Weight Number</u>	<u>Tolerance</u>	<u>Accept/Reject</u>
50 lb	B-C-11	1	F	A
50 lb	B-C-12	1	F	A
50 lb	B-C-2	1	F	A
50 lb	B-C-3	1	F	A
50 lb	B-C-4	1	F	A
50 lb	B-C-5	1	F	A
50 lb	B-C-6	1	F	A
50 lb	B-C-7	1	F	A
50 lb	B-C-8	1	F	A
50 lb	B-C-9	1	F	A
2 kg	K2	1	F	A
1000 lb	OA20	1	F	A
25 lb	OPI-D1	1	F	A
1 kg	WM-2-89-3	1	F	A
500 g	WM-2-89-3	2	F	A
200 g	WM-2-89-3	3	F	A
200 g*	WM-2-89-3	4	F	A
100 g	WM-2-89-3	5	F	A
50 g	WM-2-89-3	6	F	A
20 g	WM-2-89-3	7	F	A
20 g*	WM-2-89-3	8	F	A
10 g	WM-2-89-3	9	F	A
5 g	WM-2-89-3	10	F	A
2 g*	WM-2-89-3	12	F	A
1 g	WM-2-89-3	13	F	A
500 mg	WM-2-89-3	14	F	A
200 mg	WM-2-89-3	15	F	A
200 mg*	WM-2-89-3	16	F	A
100 mg	WM-2-89-3	17	F	A
5 lb	WM-2B86	1	F	A
5 lb	WM-2B86	2	F	A
5 lb	WM-2B86	3	F	A
5 lb	WM-2B86	4	F	A
5 lb	WM-2B86	5	F	A
1 lb	WM-2B86	6	F	A
1 lb	WM-2B86	7	F	A

Submitted By: Stoney Brooks

Electronic Copy Page 3

Denomination	Serial Number	Weight Number	Tolerance	Accept/Reject
1 lb	WM-2B86	8	F	A
1 lb	WM-2B86	9	F	A
1 lb	WM-2B86	10	F	A
8 oz	WM-2B86	11	F	A
4 oz	WM-2B86	12	F	A
4 oz	WM-2B86	13	F	A
2 oz	WM-2B86	14	F	A
1 oz	WM-2B86	15	F	A
0.5 oz	WM-2B86	16	F	A
0.25 oz	WM-2B86	17	F	A
0.125 oz	WM-2B86	18	F	A
50 lb	WM50-12	1	F	A
50 lb	WM50-16	1	F	A
50 lb	WM50-52	1	F	A
50 lb	WM50-53	1	F	A
50 lb	WM50-7	1	F	A
25 lb	WM-D15	1	F	A
25 lb	WM-D21	1	F	A
25 lb	WM-D22	1	F	A
25 lb	WM-D23	1	F	A
25 lb	WM-D24	1	F	A
25 lb	WM-D25	1	F	A
25 lb	WM-D26	1	F	A
25 lb	WM-D28	1	F	A
25 lb	WM-D29	1	F	A
25 lb	WM-D44	1	F	A
25 lb	WM-OPI-D3	1	F	A

These field standards have been compared with our reference standards and found to be within the tolerances for their class, or condemned, as prescribed by the National Institute of Standards and Technology. These field standards also meet the required specifications.

The reference standards used are traceable to the National Institute of Standards and Technology (NIST). Our traceability is maintained through control charts, regional inter-comparisons and periodic recertifications. This certificate is issued by the Nebraska Department of Agriculture / Division of Weights and Measures. Reproduction of this certificate must include the entire document (including all attached papers).

See 'Test Results' attached

See report/letter attached



Craig E. Olsen

Nebraska State Metrologist

**Nebraska Standards Laboratory**  
**3721 West Cuming Street**  
**Lincoln, NE 68524**  
**(402) 471-2087**

**SOP 8 - Modified Substitution Test Results**

*Form No.: NE04LF9-R00*

Agency ID: 35  
 Agency: Food Safety & Consumer Protection  
 Area 35  
 , -

NE Cert No: 20120130  
 Date: 11/5/2012  
 P.O. No:  
 Observer: CEO

Submitted By: Stoney Brooks

Electronic Copy

Page 1 of 4

Serial Number	Wt. No.	Retest No.	Denom.	Comparator used	Standard Used	Error of test weight	Tolerance	Uncertainty	Accepted or Rejected
A5C-1	1	0	50 lb	KA30-3_50lb	NSL-50-1	1.8769199 g	2.3 g	0.0496968 g	A
A5C-11	1	0	50 lb	KA30-3_50lb	NSL-50-1	1.4869199 g	2.3 g	0.0496968 g	A
A5C-4	1	0	50 lb	KA30-3_50lb	NSL-50-1	-2.2530801 g	2.3 g	0.0496968 g	A
A5C-4	1	1	50 lb	KA30-3_50lb	NSL-50-1	0.1069199 g	2.3 g	0.0496968 g	A
B10	1	0	1000 lb	XP604KM_1K lb	C24	0.8792599 g	45 g	2.0949073 g	A
B11	1	0	1000 lb	XP604KM_1K lb	C24	-31.8207401 g	45 g	2.0949073 g	A
B12	1	0	1000 lb	XP604KM_1K lb	C24	-14.8207401 g	45 g	2.0949073 g	A
B13	1	0	1000 lb	XP604KM_1K lb	C24	-29.2207401 g	45 g	2.0949073 g	A
B14	1	0	1000 lb	XP604KM_1K lb	C24	-43.6207401 g	45 g	2.0949073 g	A
B17	1	0	1000 lb	XP604KM_1K lb	C24	-13.8207401 g	45 g	2.0949073 g	A
B18	1	0	1000 lb	XP604KM_1K lb	C24	2.7792599 g	45 g	2.0949073 g	A
B19	1	0	1000 lb	XP604KM_1K lb	C24	3.9792599 g	45 g	2.0949073 g	A
B2	1	0	1000 lb	XP604KM_1K lb	C24	-7.0207401 g	45 g	2.0949073 g	A
B20	1	0	1000 lb	XP604KM_1K lb	C24	-7.3207401 g	45 g	2.0949073 g	A
B21	1	0	1000 lb	XP604KM_1K lb	C24	1.6792599 g	45 g	2.0949073 g	A
B22	1	0	1000 lb	XP604KM_1K lb	C24	-14.4207401 g	45 g	2.0949073 g	A
B23	1	0	1000 lb	XP604KM_1K lb	C24	-6.5207401 g	45 g	2.0949073 g	A
B3	1	0	1000 lb	XP604KM_1K lb	C24	-3.1207401 g	45 g	2.0949073 g	A
B4	1	0	1000 lb	XP604KM_1K lb	C24	-8.6207401 g	45 g	2.0949073 g	A
B5	1	0	1000 lb	XP604KM_1K lb	C24	17.3792599 g	45 g	2.0949073 g	A
B6	1	0	1000 lb	XP604KM_1K lb	C24	2.8792599 g	45 g	2.0949073 g	A
B7	1	0	1000 lb	XP604KM_1K lb	C24	14.0792599 g	45 g	2.0949073 g	A
B8	1	0	1000 lb	XP604KM_1K lb	C24	16.5792599 g	45 g	2.0949073 g	A
B9	1	0	1000 lb	XP604KM_1K lb	C24	27.5792599 g	45 g	2.0949073 g	A
B-C-1	1	0	50 lb	KA30-3_50lb	NSL-50-1	0.3069199 g	2.3 g	0.0496968 g	A
B-C-10	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.6080801 g	2.3 g	0.0496968 g	A

**Retest number - 0 indicates "as received" error / 1 indicates "adjusted to" error**

**The uncertainty of measurement is determined using the uncertainty of the reference standard(s), standard deviation of the process, environmental factors and bias using a RSS calculation. Uncertainty values are calculated using a K factor of 2.**

**"As received" values for weights may not be a true "as found" value because these weights could have been cleaned and/or repainted prior to submission to this lab. Due to handling, storage and/or environmental factors; the reported values are valid only while test items are within our control. No calibration period or interval is given with this report. We can not guarantee accuracy of results beyond the controls of our laboratory.**

Agency ID: 35  
 Agency: Food Safety & Consumer Protection  
 Area 35

NE Cert No: 20120130  
 Date: 11/5/2012  
 P.O. No:  
 Observer: CEO

Submitted By: Stoney Brooks

Electronic Copy

Page 2 of 4

Serial Number	Wt. No.	Retest No.	Denom.	Comparator used	Standard Used	Error of test weight	Tolerance	Uncertainty	Accepted or Rejected
B-C-11	1	0	50 lb	KA30-3_50lb	NSL-50-1	0.1019199 g	2.3 g	0.0496968 g	A
B-C-12	1	0	50 lb	KA30-3_50lb	NSL-50-1	1.3369199 g	2.3 g	0.0496968 g	A
B-C-2	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.2680801 g	2.3 g	0.0496968 g	A
B-C-3	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.3330801 g	2.3 g	0.0496968 g	A
B-C-4	1	0	50 lb	KA30-3_50lb	NSL-50-1	2.3219199 g	2.3 g	0.0496968 g	R
B-C-4	1	1	50 lb	KA30-3_50lb	NSL-50-1	1.2119199 g	2.3 g	0.0496968 g	A
B-C-5	1	0	50 lb	KA30-3_50lb	NSL-50-1	-3.4580801 g	2.3 g	0.0496968 g	R
B-C-5	1	1	50 lb	KA30-3_50lb	NSL-50-1	-0.7530801 g	2.3 g	0.0496968 g	A
B-C-6	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.4680801 g	2.3 g	0.0496968 g	A
B-C-7	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.5130801 g	2.3 g	0.0496968 g	A
B-C-8	1	0	50 lb	KA30-3_50lb	NSL-50-1	-1.9580801 g	2.3 g	0.0496968 g	A
B-C-8	1	1	50 lb	KA30-3_50lb	NSL-50-1	0.0719199 g	2.3 g	0.0496968 g	A
B-C-9	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.9230801 g	2.3 g	0.0496968 g	A
K2	1	0	2 kg	CC10000S_2kg	2kgd	-0.1787774 g	0.2 g	0.0038731 g	A
OA20	1	0	1000 lb	XP604KM_1K lb	C24	16.6792599 g	45 g	2.0949073 g	A
OPI-D1	1	0	25 lb	KA30-3_25lb	NSL-25-1	-0.2052207 g	1.1 g	0.0365186 g	A
WM-2-89-3	1	0	1 kg	CC1201_1 kg	1kgd	0.0513716 g	0.1 g	0.0003588 g	A
WM-2-89-3	2	0	500 g	CC1201_1 kg	500gd	0.0388294 g	0.07 g	0.0003542 g	A
WM-2-89-3	3	0	200 g	CC1201_1 kg	200gd	0.0223214 g	0.04 g	0.0003530 g	A
WM-2-89-3	4	0	200 g*	CC1201_1 kg	200gd	0.0209214 g	0.04 g	0.0003530 g	A
WM-2-89-3	5	0	100 g	AT106_100g	100gv	-0.0006285 g	0.02 g	0.0000569 g	A
WM-2-89-3	6	0	50 g	AT106_50g	50gv	0.0059558 g	0.01 g	0.0000192 g	A
WM-2-89-3	7	0	20 g	AT106_20g	20gv	-0.0000484 g	0.004 g	0.0000089 g	A
WM-2-89-3	8	0	20 g*	AT106_20g	20gv	-0.0010404 g	0.004 g	0.0000089 g	A
WM-2-89-3	9	0	10 g	AT106_10g	10gv	0.0007662 g	0.002 g	0.0000105 g	A
WM-2-89-3	10	0	5 g	CCE6_5 gm	5gv	-0.0005678 g	0.0015 g	0.0000083 g	A
WM-2-89-3	12	0	2 g*	CCE6_5 gm	2gv	0.0005852 g	0.0011 g	0.0000044 g	A
WM-2-89-3	13	0	1 g	CCE6_1 gm	1gv	0.0004852 g	0.0009 g	0.0000031 g	A
WM-2-89-3	14	0	500 mg	CCE6_1 gm	500mgv	0.0004846 g	0.00072 g	0.0000030 g	A
WM-2-89-3	15	0	200 mg	CCE6_1 gm	200mgv	0.0003067 g	0.00054 g	0.0000029 g	A
WM-2-89-3	16	0	200 mg*	CCE6_1 gm	200mgv	0.0003578 g	0.00054 g	0.0000029 g	A
WM-2-89-3	17	0	100 mg	CCE6_100 mg	100mgv	-0.0001081 g	0.00043 g	0.0000044 g	A
WM-2B86	1	0	5 lb	CC10000S_5lb	NSL-5-1	-0.0241534 g	0.23 g	0.0016377 g	A

*Retest number - 0 indicates "as received" error / 1 indicates "adjusted to" error*

*The uncertainty of measurement is determined using the uncertainty of the reference standard(s), standard deviation of the process, environmental factors and bias using a RSS calculation. Uncertainty values are calculated using a K factor of 2.*

*"As received" values for weights may not be a true "as found" value because these weights could have been cleaned and/or repainted prior to submission to this lab. Due to handling, storage and/or environmental factors; the reported values are valid only while test items are within our control. No calibration period or interval is given with this report. We can not guarantee accuracy of results beyond the controls of our laboratory.*

Agency ID: 35  
 Agency: Food Safety & Consumer Protection  
 Area 35

NE Cert No: 20120130  
 Date: 11/5/2012  
 P.O. No:  
 Observer: CEO

Submitted By: Stoney Brooks

Electronic Copy

Page 3 of 4

Serial Number	Wt. No.	Retest No.	Denom.	Comparator used	Standard Used	Error of test weight	Tolerance	Uncertainty	Accepted or Rejected
WM-2B86	2	0	5 lb	CC10000S_5lb	NSL-5-1	-0.0050534 g	0.23 g	0.0016377 g	A
WM-2B86	3	0	5 lb	CC10000S_5lb	NSL-5-1	-0.0586534 g	0.23 g	0.0016377 g	A
WM-2B86	4	0	5 lb	CC10000S_5lb	NSL-5-1	-0.0204534 g	0.23 g	0.0016377 g	A
WM-2B86	5	0	5 lb	CC10000S_5lb	NSL-5-1	0.0059466 g	0.23 g	0.0016377 g	A
WM-2B86	6	0	1 lb	CC1201_1 lb	NSL-1-1	-0.0089445 g	0.07 g	0.0004523 g	A
WM-2B86	7	0	1 lb	CC1201_1 lb	NSL-1-1	-0.0118445 g	0.07 g	0.0004523 g	A
WM-2B86	8	0	1 lb	CC1201_1 lb	NSL-1-1	0.0099555 g	0.07 g	0.0004523 g	A
WM-2B86	9	0	1 lb	CC1201_1 lb	NSL-1-1	-0.0110445 g	0.07 g	0.0004523 g	A
WM-2B86	10	0	1 lb	CC1201_1 lb	NSL-1-1	-0.0054445 g	0.07 g	0.0004523 g	A
WM-2B86	11	0	8 oz	CC1201_8 oz	NSL-8-1	-0.0023793 g	0.045 g	0.0004066 g	A
WM-2B86	12	0	4 oz	CC1201_8 oz	NSL-Z	-0.0181279 g	0.023 g	0.0003915 g	A
WM-2B86	13	0	4 oz	CC1201_8 oz	NSL-Z	0.0091721 g	0.023 g	0.0003915 g	A
WM-2B86	14	0	2 oz	AT106_0.2 lb	NSL-Y	0.0026712 g	0.011 g	0.0000612 g	A
WM-2B86	15	0	1 oz	AT106_0.2 lb	2A9-1oz	-0.0002541 g	0.0054 g	0.0003667 g	A
WM-2B86	16	0	0.5 oz	AT106_0.05 lb	2A9-1.2oz	-0.0000481 g	0.0028 g	0.0003638 g	A
WM-2B86	17	0	0.25 oz	AT106_0.05 lb	2A9-1.4oz	-0.0001215 g	0.0017 g	0.0000269 g	A
WM-2B86	18	0	0.125 oz	AT106_0.05 lb	2A9-1.8oz	0.0001728 g	0.0013 g	0.0000224 g	A
WM50-12	1	0	50 lb	KA30-3_50lb	NSL-50-1	-1.7130801 g	2.3 g	0.0496968 g	A
WM50-16	1	0	50 lb	KA30-3_50lb	NSL-50-1	1.5569199 g	2.3 g	0.0496968 g	A
WM50-52	1	0	50 lb	KA30-3_50lb	NSL-50-1	2.4619199 g	2.3 g	0.0496968 g	R
WM50-52	1	1	50 lb	KA30-3_50lb	NSL-50-1	0.1319199 g	2.3 g	0.0496968 g	A
WM50-53	1	0	50 lb	KA30-3_50lb	NSL-50-1	-0.0930801 g	2.3 g	0.0496968 g	A
WM50-7	1	0	50 lb	KA30-3_50lb	NSL-50-1	-1.1980801 g	2.3 g	0.0496968 g	A
WM-D15	1	0	25 lb	KA30-3_25lb	NSL-25-1	-0.3352207 g	1.1 g	0.0365186 g	A
WM-D21	1	0	25 lb	KA30-3_25lb	NSL-25-1	0.5797793 g	1.1 g	0.0365186 g	A
WM-D22	1	0	25 lb	KA30-3_25lb	NSL-25-1	1.4547793 g	1.1 g	0.0365186 g	R
WM-D22	1	1	25 lb	KA30-3_25lb	NSL-25-1	0.2947793 g	1.1 g	0.0365186 g	A
WM-D23	1	0	25 lb	KA30-3_25lb	NSL-25-1	-0.2302207 g	1.1 g	0.0365186 g	A
WM-D24	1	0	25 lb	KA30-3_25lb	NSL-25-1	0.5247793 g	1.1 g	0.0365186 g	A
WM-D25	1	0	25 lb	KA30-3_25lb	NSL-25-1	-0.6702207 g	1.1 g	0.0365186 g	A
WM-D26	1	0	25 lb	KA30-3_25lb	NSL-25-1	0.1547793 g	1.1 g	0.0365186 g	A
WM-D28	1	0	25 lb	KA30-3_25lb	NSL-25-1	-0.8152207 g	1.1 g	0.0365186 g	A
WM-D29	1	0	25 lb	KA30-3_25lb	NSL-25-1	1.2647793 g	1.1 g	0.0365186 g	R

*Retest number - 0 indicates "as received" error / 1 indicates "adjusted to" error*

*The uncertainty of measurement is determined using the uncertainty of the reference standard(s), standard deviation of the process, environmental factors and bias using a RSS calculation. Uncertainty values are calculated using a K factor of 2.*

*"As received" values for weights may not be a true "as found" value because these weights could have been cleaned and/or repainted prior to submission to this lab. Due to handling, storage and/or environmental factors; the reported values are valid only while test items are within our control. No calibration period or interval is given with this report. We can not guarantee accuracy of results beyond the controls of our laboratory.*

Agency ID: 35  
Agency: Food Safety & Consumer Protection  
Area 35  
, -

NE Cert No: 20120130  
Date: 11/5/2012  
P.O. No:  
Observer: CEO

Submitted By: Stoney Brooks

Electronic Copy

Page 4 of 4

Serial Number	Wt. No.	Retest No.	Denom.	Comparator used	Standard Used	Error of test weight	Tolerance	Uncertainty	Accepted or Rejected
WM-D29	1	1	25 lb	KA30-3_25lb	NSL-25-1	0.0297793 g	1.1 g	0.0365186 g	A
WM-D44	1	0	25 lb	KA30-3_25lb	NSL-25-1	0.5447793 g	1.1 g	0.0365186 g	A
WM-OPI-D3	1	0	25 lb	KA30-3_25lb	NSL-25-1	0.9597793 g	1.1 g	0.0365186 g	A

*Retest number - 0 indicates "as received" error / 1 indicates "adjusted to" error*

*The uncertainty of measurement is determined using the uncertainty of the reference standard(s), standard deviation of the process, environmental factors and bias using a RSS calculation. Uncertainty values are calculated using a K factor of 2.*

*"As received" values for weights may not be a true "as found" value because these weights could have been cleaned and/or repainted prior to submission to this lab. Due to handling, storage and/or environmental factors; the reported values are valid only while test items are within our control. No calibration period or interval is given with this report. We can not guarantee accuracy of results beyond the controls of our laboratory.*