



Dave Heineman
Governor

State of Nebraska

DEPARTMENT OF AGRICULTURE
GREGORY A. IBACH
Director

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524

CERTIFICATION FOR STANDARDS OF MEASUREMENT

Agency: Dept of Ag. - Weights & Measures
17100 SW 14th Street
Martell, NE 68404-6002

Date: 9/7/2010
Agency ID: 70
NE Cert No: 20100118
P.O. No:

Submitted By: Scott Arner

Electronic Copy Page 1

Denomination	Serial Number	Weight Number	Tolerance Class
100 gal	18969	1	VL
20 gal	88220	1	VP
100 gal	8851397	1	VL
103 gal	A-4-L6998	1	VP

These field standards have been compared with our reference standards and found to be within the tolerances for their class, or condemned, as prescribed by the National Institute of Standards and Technology. These field standards also meet the required specifications.

The reference standards used are traceable to the National Institute of Standards and Technology (NIST). Our traceability is maintained through control charts, regional inter-comparisons and periodic recertifications. This certificate is issued by the Nebraska Department of Agriculture / Division of Weights and Measures. Reproduction of this certificate must include the entire document (including all attached papers).

- See 'Test Results' attached
- See report/letter attached

Craig E. Olsen
Nebraska State Metrologist

Department of Agriculture, 301 Centennial Mall South
Web Site: www.agr.ne.gov

Administration
P.O. Box 94947
Lincoln, NE 68509-4947
(402) 471-2341
FAX: (402) 471-6876

Agriculture Laboratories
3703 South 14th Street
Lincoln, NE 68502-5399
(402) 471-2176
FAX: (402) 471-0091

Bureau of Animal Industry
P.O. Box 94987
Lincoln, NE 68509-4787
(402) 471-2351
FAX: (402) 471-6893

Bureau of Dairies & Foods
P.O. Box 95064
Lincoln, NE 68509-5064
(402) 471-2536
FAX: (402) 471-2759

Bureau of Plant Industry
P.O. Box 94956
Lincoln, NE 68509-4756
(402) 471-2394
FAX: (402) 471-6892

Weights and Measures
P.O. Box 94757
Lincoln, NE 68509-4757
(402) 471-4292
FAX: (402) 471-2759

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 21 - Pressure Volume (LPG) Test Results

Agency ID: 70
Agency: Dept of Ag. - Weights & Measures
 17100 SW 14th Street
 Martell, NE 68404-6002

NE Cert No: 20100118
Date: 9/7/2010
P.O. No:

Form No.: NE04LF14-R00

Submitted By: Scott Arner

Page 1 of 3

Serial No: 88220	Eq Vol/gal: 20 gal
Retest No: 0	Tolerance Class: VP
Denomination: 20 gal	Material: MS
Observer: CEO	Coef of Expansion: 0.0000186 /°F

TEST DATA

Del#	Std ID	Delivery	Mat	Coef of Exp	Water Temp	Water Density	Delta
1	5gal	4.9991	MS	0.0000265 /°F	68.9 °F	0.998098	0 gal.
2	5gal	4.9991	MS	0.0000265 /°F	68.4 °F	0.998156	0 gal.
3	5gal	4.9991	MS	0.0000265 /°F	68.1 °F	0.998191	0 gal.
4	5gal	4.9991	MS	0.0000265 /°F	67.3 °F	0.998282	0 gal.

Test Measure Information

Water Temp: 68.5 °F
Water Density: 0.998145 g/cm3

Reading @ 0 psi 0.04 gal.
Reading @ 50 psi 0.02 gal.
Reading @ 100 psi 0.01 gal.
Reading @ 150 psi 0 gal.
Reading @ 200 psi 0.02 gal.

Final Test Results

Z60 Delivery: 19.998 gal.
Prover error @ 100 psi: -0.018 gal.
Uncertainty of measurement: 0.002 gal.
Tolerance: 0.040 gal.
Accept/Reject: A

Prover Reading Corrections*

Press Corr @ 0 psi:	-0.024 gal.
Press Corr @ 50 psi:	-0.007 gal.
Press Corr @ 100 psi:	0.000 gal.
Press Corr @ 150 psi:	0.007 gal.
Press Corr @ 200 psi:	-0.016 gal.

* This table is to be used in making pressure corrections to prover readings.

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 21 - Pressure Volume (LPG) Test Results

Agency ID: 70
Agency: Dept of Ag. - Weights & Measures
 17100 SW 14th Street
 Martell, NE 68404-6002

NE Cert No: 20100118
Date: 9/7/2010
P.O. No:

Form No.: NE04LF14-R00

Submitted By: Scott Arner

Page 2 of 3

Serial No: A-4-L6998
Retest No: 0
Denomination: 103 gal
Observer: CEO

Eq Vol/gal: 103 gal
Tolerance Class: VP
Material: MS
Coef of Expansion: 0.0000186 /°F

TEST DATA

Del#	Std ID	Delivery	Mat	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	MS	0.0000265 /°F	66.8 °F	0.998338	0 gal.
2	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.
3	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.
4	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.

Test Measure Information

Water Temp: 68.5 °F
Water Density: 0.998145 g/cm3

Reading @ 0 psi 0.16 gal.
Reading @ 50 psi 0.02 gal.
Reading @ 100 psi -0.07 gal.
Reading @ 150 psi -0.15 gal.
Reading @ 200 psi -0.22 gal.

Final Test Results

Z60 Delivery: 103.009 gal.
Prover error @ 100 psi: 0.046 gal.
Uncertainty of measurement: 0.022 gal.
Tolerance: 0.050 gal.
Accept/Reject: A

Prover Reading Corrections*

Press Corr @ 0 psi:	-0.197 gal.
Press Corr @ 50 psi:	-0.073 gal.
Press Corr @ 100 psi:	0.000 gal.
Press Corr @ 150 psi:	0.063 gal.
Press Corr @ 200 psi:	0.117 gal.

* This table is to be used in making pressure corrections to prover readings.

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 21 - Pressure Volume (LPG) Test Results

Agency ID: 70
Agency: Dept of Ag. - Weights & Measures
 17100 SW 14th Street
 Martell, NE 68404-6002

NE Cert No: 20100118
Date: 9/7/2010
P.O. No:

Form No.: NE04LF14-R00

Submitted By: Scott Arner

Page 3 of 3

Serial No: A-4-L6998
Retest No: 1
Denomination: 103 gal
Observer: CEO

Eq Vol/gal: 103 gal
Tolerance Class: VP
Material: MS
Coef of Expansion: 0.0000186 /°F

TEST DATA

Del#	Std ID	Delivery	Mat	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	MS	0.0000265 /°F	66.8 °F	0.998338	0 gal.
2	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.
3	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.
4	1gal	1	MS	0.00001 /°F	66.8 °F	0.998338	0 gal.

Test Measure Information

Water Temp: 67 °F
Water Density: 0.998315 g/cm3

Reading @ 0 psi 0.18 gal.
Reading @ 50 psi 0.08 gal.
Reading @ 100 psi 0 gal.
Reading @ 150 psi -0.08 gal.
Reading @ 200 psi -0.16 gal.

Final Test Results

Z60 Delivery: 102.994 gal.
Prover error @ 100 psi: -0.039 gal.
Uncertainty of measurement: 0.022 gal.
Tolerance: 0.050 gal.
Accept/Reject: A

Prover Reading Corrections*

Press Corr @ 0 psi:	-0.147 gal.
Press Corr @ 50 psi:	-0.063 gal.
Press Corr @ 100 psi:	0.000 gal.
Press Corr @ 150 psi:	0.063 gal.
Press Corr @ 200 psi:	0.127 gal.

* This table is to be used in making pressure corrections to prover readings.

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Form No.: NE03LF12-R00

Agency: Dept of Ag. - Weights & Measures
17100 SW 14th Street
Martell, NE 68404-6002

NE Cert No: 20100118

Date: 9/7/2010

P.O. No:

Submitted By: Scott Arner

Page 1 of 4

Serial #: 18969
Retest #: 0
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	67.9 °F	0.998214	0 in3

Test Measure Information

Water Temperature: 67.1 °F
Water Density: 0.998214 g/cm³
Scale Reading: -1 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 99.9846 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.
Accepted/Rejected: A

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Form No.: NE03LF12-R00

Agency: Dept of Ag. - Weights & Measures
17100 SW 14th Street
Martell, NE 68404-6002

NE Cert No: 20100118

Date: 9/7/2010

P.O. No:

Submitted By: Scott Arner

Page 2 of 4

Serial #: 18969
Retest #: 1
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	68.3 °F	0.998168	0 in3

Test Measure Information

Water Temperature: 67.4 °F
Water Density: 0.998168 g/cm³
Scale Reading: -4 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 99.9966 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.
Accepted/Rejected: A

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Form No.: NE03LF12-R00

Agency: Dept of Ag. - Weights & Measures
17100 SW 14th Street
Martell, NE 68404-6002

NE Cert No: 20100118

Date: 9/7/2010

P.O. No:

Submitted By: Scott Arner

Page 3 of 4

Serial #: 8851397
Retest #: 0
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	66.4 °F	0.998382	0 in3

Test Measure Information

Water Temperature: 65.4 °F
Water Density: 0.998382 g/cm³
Scale Reading: -5 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 100.0007 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.
Accepted/Rejected: A

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Form No.: NE03LF12-R00

Agency: Dept of Ag. - Weights & Measures
17100 SW 14th Street
Martell, NE 68404-6002

NE Cert No: 20100118

Date: 9/7/2010

P.O. No:

Submitted By: Scott Arner

Page 4 of 4

Serial #: 8851397
Retest #: 1
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	66.6 °F	0.998360	0 in3

Test Measure Information

Water Temperature: 65.6 °F
Water Density: 0.998360 g/cm³
Scale Reading: -5 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 100.0006 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.
Accepted/Rejected: A