



Dave Heineman
Governor

State of Nebraska

DEPARTMENT OF AGRICULTURE
GREGORY A. IBACH
Director

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524

CERTIFICATION FOR STANDARDS OF MEASUREMENT

Agency: Food Safety & Consumer Protection
Area 70

Date: 9/6/2011
Agency ID: 70
NE Cert No: 20110106
P.O. No:

Page 1

Submitted By: Scott Arner

Denomination	Serial Number	Weight Number	Tolerance Class
100 gal	18969	1	VL
20 gal	88220	1	VP
100 gal	8851397	1	VL
103 gal	A-4-L6998	1	VP

These field standards have been compared with our reference standards and found to be within the tolerances for their class, or condemned, as prescribed by the National Institute of Standards and Technology. These field standards also meet the required specifications.

The reference standards used are traceable to the National Institute of Standards and Technology (NIST). Our traceability is maintained through control charts, regional inter-comparisons and periodic recertifications. This certificate is issued by the Nebraska Department of Agriculture / Division of Weights and Measures.

See 'Test Results' attached

See report/letter attached

Craig E. Olsen

Nebraska State Metrologist

Department of Agriculture, 301 Centennial Mall South
Web Site: www.agr.ne.gov

Administration
P.O. Box 94947
Lincoln, NE 68509-4947
(402) 471-2341
FAX: (402) 471-6876

Agriculture Laboratories
3703 South 14th Street
Lincoln, NE 68502-5399
(402) 471-2176
FAX: (402) 471-0091

Bureau of Animal Industry
P.O. Box 94987
Lincoln, NE 68509-4787
(402) 471-2351
FAX: (402) 471-6893

Bureau of Dairies & Foods
P.O. Box 95064
Lincoln, NE 68509-5064
(402) 471-2536
FAX: (402) 471-2759

Bureau of Plant Industry
P.O. Box 94956
Lincoln, NE 68509-4756
(402) 471-2394
FAX: (402) 471-6892

Weights and Measures
P.O. Box 94757
Lincoln, NE 68509-4757
(402) 471-4292
FAX: (402) 471-2759

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Agency: Food Safety & Consumer Protection
Area 70

Form No.: NE03LF12-R00

NE Cert No: 20110106

Date: 9/6/2011

P.O. No:

Submitted By: Scott Arner

Page 1 of 4

Serial #: 18969
Retest #: 0
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	68.7 °F	0.998121	0 in3

Test Measure Information

Water Temperature: 68.2 °F
Water Density: 0.998121 g/cm³
Scale Reading: -2.5 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 99.9935 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.

Accepted/Rejected: A

1 cubic meter = 264.172 gallons

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Agency: Food Safety & Consumer Protection
Area 70

Form No.: NE03LF12-R00

NE Cert No: 20110106

Date: 9/6/2011

P.O. No:

Submitted By: Scott Arner

Page 2 of 4

Serial #: 18969
Retest #: 1
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	68.8 °F	0.998110	0 in3

Test Measure Information

Water Temperature: 68.4 °F
Water Density: 0.998110 g/cm³
Scale Reading: -3 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 99.9966 gal.
Tolerance: 0.0519 gal.
Uncertainty: 0.0223 gal.

Accepted/Rejected: A

1 cubic meter = 264.172 gallons

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Agency: Food Safety & Consumer Protection
Area 70

Form No.: NE03LF12-R00

NE Cert No: 20110106

Date: 9/6/2011

P.O. No:

Submitted By: Scott Arner

Page 3 of 4

Serial #: 8851397
Retest #: 0
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	68.7 °F	0.998121	0 in3

Test Measure Information

Water Temperature: 68.5 °F
Water Density: 0.998121 g/cm³
Scale Reading: -5 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 100.0070 gal.

Tolerance: 0.0519 gal.

Uncertainty: 0.0223 gal.

Accepted/Rejected: A

1 cubic meter = 264.172 gallons

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 19 - Volume Transfer Test Results

Agency ID: 70

Agency: Food Safety & Consumer Protection
Area 70

Form No.: NE03LF12-R00

NE Cert No: 20110106

Date: 9/6/2011

P.O. No:

Submitted By: Scott Arner

Page 4 of 4

Serial #: 8851397
Retest #: 1
Denomination: 100 gal
Observer: CEO

Eq Vol/cu.in.: 23100 cu.in.
Tolerance Class: VL
Material: SS
Coef of Expansion: 0.0000265 °F

Test Data

Del#	Std ID	Delivery	Material	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	SS	0.0000265 /°F	68.7 °F	0.998121	0 in3

Test Measure Information

Water Temperature: 68.6 °F
Water Density: 0.998121 g/cm³
Scale Reading: -5 cu.in.

Test Results

Test Measure Volume @ 60° F at zero scale reading: 100.0079 gal.

Tolerance: 0.0519 gal.

Uncertainty: 0.0223 gal.

Accepted/Rejected: A

1 cubic meter = 264.172 gallons

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 21 - Pressure Volume (LPG) Test Results

Agency ID: 70
Agency: Food Safety & Consumer Protection
 Area 70

NE Cert No: 20110106
Date: 9/6/2011
P.O. No:

Form No.: NE04LF14-R00

Submitted By: Scott Arner

Page 1 of 2

Serial No: 88220
Retest No: 0
Denomination: 20 gal
Observer: CEO

Eq Vol/gal: 20 gal
Tolerance Class: VP
Material: MS
Coef of Expansion: 0.0000186 /°F

TEST DATA

Del#	Std ID	Delivery	Mat	Coef of Exp	Water Temp	Water Density	Delta
1	5gal	4.9991	MS	0.0000265 /°F	68.4 °F	0.998156	0 gal.
2	5gal	4.9991	MS	0.0000265 /°F	68.7 °F	0.998121	0 gal.
3	5gal	4.9991	MS	0.0000265 /°F	67.9 °F	0.998214	0 gal.
4	5gal	4.9991	MS	0.0000265 /°F	67.2 °F	0.998293	0 gal.

Test Measure Information

Water Temp: 69 °F
Water Density: 0.998086 g/cm³

Reading @ 0 psi 0.05 gal.
Reading @ 50 psi 0.02 gal.
Reading @ 100 psi 0.01 gal.
Reading @ 150 psi -0.01 gal.
Reading @ 200 psi -0.02 gal.

Final Test Results

Z60 Delivery: 20.000 gal.
Prover error @ 100 psi: -0.017 gal.
Uncertainty of measurement: 0.002 gal.
Tolerance: 0.040 gal.
Accept/Reject: A

Prover Reading Corrections*

Press Corr @ 0 psi:	-0.034 gal.
Press Corr @ 50 psi:	-0.007 gal.
Press Corr @ 100 psi:	0.000 gal.
Press Corr @ 150 psi:	0.017 gal.
Press Corr @ 200 psi:	0.024 gal.

1 cubic meter = 264.172 gallons

* This table is to be used in making pressure corrections to prover readings.

Nebraska Standards Laboratory
3721 West Cuming Street
Lincoln, NE 68524
(402) 471-2087

SOP 21 - Pressure Volume (LPG) Test Results

Agency ID: 70
Agency: Food Safety & Consumer Protection
 Area 70

NE Cert No: 20110106
Date: 9/6/2011
P.O. No:

Form No.: NE04LF14-R00

Submitted By: Scott Arner

Page 2 of 2

Serial No: A-4-L6998	Eq Vol/gal: 103 gal
Retest No: 0	Tolerance Class: VP
Denomination: 103 gal	Material: MS
Observer: CEO	Coef of Expansion: 0.0000186 /°F

TEST DATA

Del#	Std ID	Delivery	Mat	Coef of Exp	Water Temp	Water Density	Delta
1	100gal	99.9872	MS	0.0000265 /°F	68.6 °F	0.998133	0 gal.
2	1gal	1	MS	0.00001 /°F	68.6 °F	0.998133	0 gal.
3	1gal	1	MS	0.00001 /°F	68.6 °F	0.998133	0 gal.
4	1gal	1	MS	0.00001 /°F	68.6 °F	0.998133	0 gal.

Test Measure Information

Water Temp: 69 °F
Water Density: 0.998086 g/cm³

Reading @ 0 psi 0.32 gal.
Reading @ 50 psi 0.17 gal.
Reading @ 100 psi 0.08 gal.
Reading @ 150 psi -0.01 gal.
Reading @ 200 psi -0.1 gal.

Final Test Results

Z60 Delivery: 102.998 gal.
Prover error @ 100 psi: -0.115 gal.
Uncertainty of measurement: 0.022 gal.
Tolerance: 0.200 gal.
Accept/Reject: A

Prover Reading Corrections*

Press Corr @ 0 psi:	-0.207 gal.
Press Corr @ 50 psi:	-0.073 gal.
Press Corr @ 100 psi:	0.000 gal.
Press Corr @ 150 psi:	0.073 gal.
Press Corr @ 200 psi:	0.147 gal.

1 cubic meter = 264.172 gallons

* This table is to be used in making pressure corrections to prover readings.